

TeltoHeart UKCA



We, UAB TELTONIKA TELEMEDIC, declare under our sole responsibility that the above described product is in conformity with the relevant Community harmonization: Union Directive RER 2017 (SI 2017/1206).

Description

The UKCA (UK Conformity Assessed) marking **is a new UK product marking that is used for goods being placed on the market in Great Britain (England, Wales and Scotland)**. It covers most goods which previously required the CE marking , known as 'new approach' goods. The UKCA marking came into effect on 1 January 2021.

Standards applied

The conformity with the essential requirements has been demonstrated against the following harmonized standards and other technical specifications:

| Standard | Article | Report No. |
|---|--|------------------------------|
| EN IEC 62311:2020 EN 62368-1:2014+A11:2017 | Health and safety - Article 3.1(a) | T251-0192/23 C223-0009/23 |
| EN 303 446-1 V1.2.1 | Electromagnetic compatibility - Article 3.1(b) | T251-0197/23 |
| EN300 328 V2.2.2 | Efficient use of radio spectrum - Article 3.2 | T251-0174/23 |

Therefore  is placed on the product.

Disclaimer:

Test reports that are referenced in declarations and certificates can be provided upon request. For the request to be approved, the recipient of the test reports should be a certification authority or certified test house. The recipients will be asked to sign a non-disclosure agreement (NDA).

Attachments



UKCA Declaration of conformity, TeltoHeart
page 1



UKCA Declaration of conformity, TeltoHeart
page 2

You can find PDF version of the UKCA declaration of conformity [here](#).